

August 4, 2015

**Dräger Medical Verification of Meru Wireless MC4200 LAN Controller and Meru AP320 Access Points with Wireless Network for OneNet<sup>®</sup> (shared Dräger Infinity network) M300 Installations**

Dear Madam/Sir,

Dräger Medical is pleased to announce that the verification of the Meru MC4200 Controller and AP320 Access Points with 7.0-8-0 firmware for M300 installations. These components have been extended to the Infinity OneNet approved network equipment checklist. Infinity OneNet is the award-winning solution introduced by Dräger Medical that enables the integration of Infinity patient monitors as part of a single hospital wide network.

The following network components are verified for use with the Dräger Medical OneNet network:

<b>WLAN Controller</b>	<b>Access Points</b>	<b>Tested Software Version</b>
Meru MC4200	Meru AP320	7.0-8-0

The goal of this verification initiative is to investigate situations where customers want to operate Dräger and non-Dräger clients on shared IEEE 802.11 a/b/g/n access points. The above referenced Meru wireless controller and access points passed the simulated tests where two access points were loaded 12 Infinity M300 monitors and 32 simulated wired patient monitors as well as competing traffic.

Please be advised it remains the customer's responsibility to actively monitor potential sources of wireless interference as well as changes in the wireless traffic load.

Yours Sincerely,



Angela Noble  
Product Manager, Infinity OneNet  
KAF Monitoring, Systems & IT  
Tel: +1 978 379-8514  
[angela.noble@draeger.com](mailto:angela.noble@draeger.com)

**Draeger Medical Systems, Inc.**

6 Tech Drive  
Andover, MA 01810 USA  
Phone: (978) 379-8000  
[www.draeger.com](http://www.draeger.com)